

U2B SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: C	Effective: 5/24/2010	Page 1 of 1	Number: LS-SCL- 0066
Prepared by: Philip Marino		Reviewed by: P. Marino	Approved by: B. Chmiel

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

SAFETY - VACUUM

1. The G1 vacuum is 4×10^{-9} Torr or less.

SAFETY – VISIBLE LIGHT HAZARD

2. Any viewport capable of transmitting the visible part of the beam must be covered.

SAFETY – SHIELDING AND RADIATION

3. No additional passive shielding is required since the beam is extracted vertically behind the existing shielding.
4. No one should remain near the front end during injection.

OPCO / Beamline Rep.: _____ Date: _____